

# JT Test after Django upgrade

## DESCRIPTION

JT Test after Django upgrade

## SPECIFICATIONS

Processing time	30–50 min
Number of places in stand	16
Processing capability	1–12 samples per run
Tube's volume	1.5 - 2 ml
Extraction technology	magnetic particle separation technology
Tube's manufacturer	Eppendorf or equivalent
Protocol	on board, preinstalled protocols
Protocol input	barcode scanner
Sample volume	100–1000 µl
Elution volume	50–300 µl
Connection to PC	not required
Display	LCD (20×4)
Certification	CE IVD
Nominal operating voltage	110–240 V, 50/60 Hz
Dimensions (W×D×H)	480×700×520 mm
Weight	43 kg

## ACCESSORIES



## CAT. NUMBER

jt-100

jt 100



#### ScienCell products

ScienCell Research Laboratories ([ScienCell](#)) is an expanding biotechnology company whose mission is the research and development of cell and cell-related products for experimental and therapeutic use. The ScienCell ...

[read more](#)



**Bio PP-4**  
BS-010116-AK  
platform

Flat platform with non-slip silicone mat for Petri dishes, culture flasks, agglutination.